## One touch and off it goes ...



# QUICKCHECK webline ®

Portable Constancy Check Device



## **Completely Automatic Measurement**

## Set one time, use day after day:

Complete LINAC QA in an automated measurement procedure - simple, compact and wireless.

PITW QUICKCHECK webline

Siemens A1

WL: Siemens

EL 8 MeV

DOSE

FLAT

50 SYM GT

QUALITY

RATE

TIME

20"20 SYM LR

100

45

Delete

0

101,9

98,5

101.8

96.2

101.4

Calibrate

5.023 Imin

33.5 5 History No: 01/12

Setup

ACHTURIC: DIESEN GEREICH MONT RESTRATIEN

Unit

Energy

Field

SSD

Gantry

Wedge

Last

MU

OTOL

BAT

RDY

## **Routine and Time-Saving LINAC QA**

FFF Compensator (X6, X10) available

to perform LINAC QA of FFF beams just as you

are used to it.

# **OUICKCHECK** webline ®

Intuitive operability
Ready for use immediately every day
Track-it data management

dynamic and static wedges according to pre-defined task lists

## SET ONE TIME, USE DAY AFTER DAY

A task list is created for each linear accelerator in radiotherapy using the WorklistGenerator: The software makes it possible to define all relevant test parameters for comprehensive quality control of accelerators and apply them in a constant, completely automatic measurement routine. The task lists are transmitted to the measurement equipment and stored there. Thereafter, QUICKCHECK<sup>webline</sup> is ready for daily constancy checks without the need of additional settings.

### 1. Aligning and irradiating

QUICKCHECK<sup>webline</sup> is placed under the linear accelerator with little work and aligned using the room laser with help of engraved lines. Thanks to a convenient auto-start and auto-stop function, one measurement after another can be performed without further intervention. For example, you can irradiate various photon and electron energies one after another. The data are stored in the device.

### 2. Displaying and evaluating

The task list for the respective linear accelerator is worked through completely, and the corresponding test status is calculated automatically. QUICKCHECK<sup>live</sup> displays instantially the measurement results. Measurement tasks can be skipped or repeated directly from outside the treatment room. QUICKCHECKwebline manages task lists for any number of accelerators; all available LINACS can be checked with only one measurement device, and the measurement values of several months can be stored in the device.

### 3. Analyzing and archiving

The data stored in QUICKCHECK webline can be transferred to a PC via RS232, USB or LAN at any time. QUICKCHECK software supports fast long-term analysis. Data transfer to Track-it offers to manage all QA data on one single platform and to share it fast and effectively across your organization. Track-it features automated completition of reports by using predefined or custom protocol templates, e.g., AAPM TG-142 Daily QA. An optional QUICKCHECKwebline docking station, which can also be mounted on the wall, serves for data transmission as well as for charging the batteries.

New Open line what Derman Prot Carters	PIV
Walter	frankes
	James Hell         James H

## Robust ionization chambers

Test device for constancy check

0.1/0.2 cm<sup>3</sup> per chamber

## **Specifications**

Measurement volume:

Produ	ct:
-------	-----

	of medical linear accelerators
Detector type:	Vented ionization chambers,
	air density compensated
Number of detectors:	13
Measured quantities:	Dose, dose rate, irradiation time,
	temperature and air pressure
Nominal useful energy range:	(4 25) MV photons
	(4 10) MV photons with
	FFF Compensator
	(4 25) MeV electrons
Nominal range of dose rate:	(0.5 10) Gy/min
	(1 25) Gy/min with
	FFF Compensator
Response:	3.4 nC/Gy (typical)

## Suitable for field sizes from 10 x 10 cm<sup>2</sup>

Field sizes: Temperature range: Humidity range: Air pressure range: Display: Interfaces: Power supply: Outer dimensions: Weight:

(10 x 10) cm<sup>2</sup>, (20 x 20) cm<sup>2</sup> (10 ... 40) °C, (50 ... 95) °F (10 ... 80) %, max. 20 g/m<sup>3</sup> (540 ... 1060) hPa TFT color display 70 mm x 53 mm RS232, USB und LAN (TCP/IP) 4 rechargeable batteries AA (NiMH) 380 mm x 254 mm x 67 mm, 262 mm x 262 mm x 37.5 mm Compensator X6, 262 mm x 262 mm x 58.5 mm Compensator X10 Measurement device 5.5 kg, Docking station 1.7 kg, Compensator X6 2.1 kg, Compensator X10 2.9 kg

## ALIGNING AND IRRADIATING



Integrated energy check for photons and electrons

User-

## Easy to operate

- Compact format and slight weight
- Detectors and display in one device
- Automatic air density correction
- Wireless operation
- Suitable gantry holding device available for measurements under various gantry angles

## DISPLAYING AND EVALUATING



defined check parameters

Comprehensive network capability with USB, RS 232 and LAN

## Completely automatic and convenient

- Task lists defined once in advance
- Trigger-controlled measurement procedure with auto-start, auto-stop, auto-standby and auto-shutdown
- Immediate display of measurement results
- Skip or repeat measurement tasks from outside the treatment room

## ANALYZING AND ARCHIVING





interfaces for PC or network connection

Data management with Track-it

## Versatile and transparent

- Individually configurable evaluation
- Export analyzed data from your QUICKCHECK<sup>webline</sup> to Track-it with the click of a single button
- ▶ View and manage all your QA reports in the Track-it Dashboard
- Access QA data from multiple sources and sites with a standard web browser

### **Dosimetry Pioneers since 1922.**

PTW is a global market leader for dosimetry and quality control solutions in radiation medicine, serving the needs of medical radiation experts in more than 160 countries worldwide. Starting with the famous Hammer dosemeter in 1922, the German manufacturer is one of the pioneers in medical radiation measurement, known for its unparalleled quality and precision.

For PTW, making medical radiation safer is both a passion and lifetime commitment. The family-run high-tech company operates one of the oldest and largest accredited calibration laboratories in the field of ionizing radiation and established THE DOSIMETRY SCHOOL to promote the exchange of knowledge in clinical dosimetry.

For more information on PTW products visit www.ptw.de or contact your local PTW representative:

### Headquarters

### PTW-Freiburg Physikalisch-Technische Werkstätten Dr. Pychlau GmbH Lörracher Straße 7 79115 Freiburg · Germany Phone +49 761 49055-0 Fax +49 761 49055-70 info@ptw.de

#### PTW-UK Ltd.

www.ptw.de

Old School House Station Road East Grantham Lincolnshire NG31 6HX · United Kingdom Phone: (+44) 1476 577503 Fax: (+44) 1476 577503 sales@ptw-uk.com www.ptw-uk.com

#### **PTW-France SARL**

41 Chemin de la Cerisaie 91620 La Ville du Bois · France Phone +33 1 64 49 98 58 Fax +33 1 69 01 59 32 info@ptw-france.com www.ptw-france.com

#### PTW Dosimetría Iberia S. L.

Calle Profesor Beltrán Báguena nº 4 - 312E 46009 Valencia · Spain Phone +34 96 346 2854 Fax +34 96 321 2140 info@ptwdi.es www.ptwdi.es

### **PTW-New York Corporation**

140 58th Street Suite # 5H3 Brooklyn, New York 11220 · USA Phone (1-516) 827 3181 Fax (1-516) 827 3184 ptw@ptwny.com www.ptwny.com

### PTW-Latin America

Av. Evandro Lins e Silva 840 Sala 2018 · Barra da Tijuca 22631-470 Rio de Janeiro-RJ · Brazil Phone +55 21 2178 2188 Fax +55 21 2429 6234 info@ptw.com.br www.ptw.com.br

#### PTW-Asia Pacific Ltd.

Workshop I on 11<sup>th</sup> Floor Valiant Industrial Centre Nos. 2-12 Au Pui Wan Street Fo Tan, New Territories · Hong Kong Phone +852 2369 9234 Fax +852 2369 9235 info@ptw-asiapacific.com www.ptw-asiapacific.com

#### PTW-Beijing Ltd

Room 712, JinYiYe Building No. 2 Sheng Gu Zhong Lu ChaoYang District 100029 Beijing · P.R. of China Phone +86 10 6443 0746 Fax +86 10 6442 7804 info@ptw-beijing.com www.ptw-beijing.com

#### PTW Dosimetry India Pvt. Ltd.

ACE Towers, 2<sup>nd</sup> Floor 73/75 Dr Radhakrishnan Road · Mylapore Chennai 600004 · India Phone +91 44 42079999 Fax +91 44 42072299 info@ptw-india.in www.ptw-india.in

©PTW. All Rights Reserved. Due to continuous innovation, product specifications are subject to change without prior notice. Printing errors and omissions excepted. All trademarks mentioned in this document are the property of their respective owners. D871.139.00/02 2019-01